

## InnoVet Technique Chart 60Hz

NOTE: 400 speed (except where noted) - automatic processing @ 94 degrees.

### InnoVet Technique Chart 60Hz - 400 speed (except where noted) - automatic processing @ 94 degrees.

Head VD		Table Top			Table w/Grid						
cm		4	6	8	10	12	14	16	18	20	22
kVp		56	60	64	68	72	76	80	84	88	92
mA		100	100	100	100	100	100	100	100	100	100
time		1/20	1/20	1/10	1/10	1/10	1/10	1/10	1/10	1/10	1/10
mAs		5	5	10	10	10	10	10	10	10	10

  

Head Lateral		Table Top			Table w/Grid						
cm		4	6	8	10	12	14	16	18	20	22
kVp		56	60	64	68	72	76	80	84	88	92
mA		100	100	100	100	100	100	100	100	100	100
time		1/20	1/20	1/10	1/10	1/10	1/10	1/10	1/10	1/10	1/10
mAs		5	5	10	10	10	10	10	10	10	10

  

Thorax VD		Table Top		Table w/Grid							
cm		4	7	10	13	16	19	22	25	28	31
kVp		52	58	64	70	76	82	88	94	100	106
mA		300	300	300	300	300	300	300	300	300	300
time		1/120	1/120	1/60	1/60	1/60	1/60	1/60	1/60	1/60	1/60
mAs		2.4	2.4	5	5	5	5	5	5	5	5

  

Thorax Lateral		Table Top		Table w/Grid							
cm		4	7	10	13	16	19	22	25	28	31
kVp		54	60	66	72	78	84	90	96	102	108
mA		300	300	300	300	300	300	300	300	300	300
time		1/120	1/120	1/60	1/60	1/60	1/60	1/60	1/60	1/60	1/60
mAs		2.4	2.4	5	5	5	5	5	5	5	5

  

Thorax Decubitus		Table Top		Table w/Grid							
cm		4	7	10	13	16	19	22	25	28	31
kVp		54	60	66	72	78	84	90	96	102	108
mA		300	300	300	300	300	300	300	300	300	300
time		1/120	1/120	1/60	1/60	1/60	1/60	1/60	1/60	1/60	1/60
mAs		2.4	2.4	5	5	5	5	5	5	5	5

  

Abdomen VD		Table Top		Table w/Grid							
cm		4	7	10	13	16	19	22	25	28	31
kVp		52	58	62	68	74	80	86	92	98	104
mA		300	300	300	300	300	300	300	300	300	300
time		1/120	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	1/30
mAs		2.5	5	10	10	10	10	10	10	10	10

  

Abdomen Lateral		Table Top		Table w/Grid							
cm		4	7	10	13	16	19	22	25	28	31
kVp		52	58	64	70	76	82	88	94	100	106
mA		300	300	300	300	300	300	300	300	300	300
time		1/120	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	1/30
mAs		2.5	5	10	10	10	10	10	10	10	10

## InnoVet Technique Chart 60Hz

NOTE: 400 speed (except where noted) - automatic processing @ 94 degrees.

<b>Cervical VD</b>		<b>Table Top</b>			<b>Table w/Grid</b>						
cm	4	6	8	10	12	14	16	18	20	22	
kVp	56	60	64	68	72	76	80	84	88	92	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/120	1/120	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	
mAs	2.5	2.5	7.5	7.5	7.5	7.5	7.5	7.5	7.5	7.5	
<b>Cervical Lateral</b>		<b>Table Top</b>			<b>Table w/Grid</b>						
cm	4	6	8	10	12	14	16	18	20	22	
kVp	56	60	64	68	72	76	80	84	88	92	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	5	5	5	10	10	10	10	10	10	10	
<b>Cervical Soft Tissue Lateral</b>		<b>Table Top</b>			<b>Table w/Grid</b>						
cm	4	6	8	10	12	14	16	18	20	22	
kVp	56	60	64	68	72	76	80	84	88	92	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/120	1/120	1/120	1/60	1/60	1/60	1/60	1/60	1/60	1/60	
mAs	2.5	2.5	2.5	5	5	5	5	5	5	5	
<b>Thoracic VD</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	4	7	10	13	16	19	22	25	28	31	
kVp	52	58	64	70	76	82	88	94	100	106	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	5	5	10	10	10	10	10	10	10	10	
<b>Thoracic Lateral</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	4	7	10	13	16	19	22	25	28	31	
kVp	52	58	64	70	76	82	88	94	100	106	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	5	5	10	10	10	10	10	10	10	10	
<b>Lumbar VD</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	4	7	10	13	16	19	22	25	28	31	
kVp	52	58	64	70	76	82	88	94	100	106	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/120	1/120	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	
mAs	2.5	2.5	7.5	7.5	7.5	7.5	7.5	7.5	7.5	7.5	
<b>Lumbar Lateral</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	4	7	10	13	16	19	22	25	28	31	
kVp	52	58	64	70	76	82	88	94	100	106	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	5	5	10	10	10	10	10	10	10	10	

## InnoVet Technique Chart 60Hz

NOTE: 400 speed (except where noted) - automatic processing @ 94 degrees.

<b>Shoulder CC</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	6	8	10	12	14	16	18	20	22	24	
kVp	60	64	68	72	76	80	84	88	92	96	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/10	1/10	1/10	1/10	1/10	1/10	1/10	1/10	
mAs	5	5	10	10	10	10	10	10	10	10	
<b>Shoulder Lateral</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	6	8	10	12	14	16	18	20	22	24	
kVp	60	64	68	72	76	80	84	88	92	96	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/10	1/10	1/10	1/10	1/10	1/10	1/10	1/10	
mAs	5	5	10	10	10	10	10	10	10	10	
<b>Pelvis VD</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	6	8	10	12	14	16	18	20	22	24	
kVp	60	64	68	72	76	80	84	88	92	96	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	5	5	10	10	10	10	10	10	10	10	
<b>Pelvis Lateral</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	6	8	10	12	14	16	18	20	22	24	
kVp	60	64	68	72	76	80	84	88	92	96	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/40	1/40	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/20	
mAs	7.5	7.5	15	15	15	15	15	15	15	15	
<b>Pelvis Distraction</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	6	8	10	12	14	16	18	20	22	24	
kVp	60	64	68	72	76	80	84	88	92	96	
mA	300	300	300	300	300	300	300	300	300	300	
time	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	5	5	10	10	10	10	10	10	10	10	
<b>Humerus CC - 100 speed</b>		<b>Table Top</b>				<b>Table w/Grid</b>					
cm	2	4	6	8	10	12	14	16	18	20	
kVp	54	58	62	66	70	74	78	82	86	90	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	
<b>Humerus Lateral - 100 speed</b>		<b>Table Top</b>				<b>Table w/Grid</b>					
cm	2	4	6	8	10	12	14	16	18	20	
kVp	54	58	62	66	70	74	78	82	86	90	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	

## InnoVet Technique Chart 60Hz

NOTE: 400 speed (except where noted) - automatic processing @ 94 degrees.

<b>Elbow CC - 100 speed</b>					<b>Table Top</b>						<b>Table w/Grid</b>										
cm	2	4	6	8	10	12	14	16	18	20	cm	2	4	6	8	10	12	14	16	18	20
kVp	52	56	60	64	70	74	78	82	86	90	kVp	52	56	60	64	70	74	78	82	86	90
mA	100	100	100	100	100	100	100	100	100	100	mA	100	100	100	100	100	100	100	100	100	100
time	1/60	1/60	1/60	1/60	1/40	1/40	1/40	1/40	1/40	1/40	time	1/60	1/60	1/60	1/60	1/40	1/40	1/40	1/40	1/40	1/40
mAs	1.7	1.7	1.7	1.7	2.5	2.5	2.5	2.5	2.5	2.5	mAs	1.7	1.7	1.7	1.7	2.5	2.5	2.5	2.5	2.5	2.5
<b>Elbow Lateral - 100 speed</b>					<b>Table Top</b>						<b>Table w/Grid</b>										
cm	2	4	6	8	10	12	14	16	18	20	cm	2	4	6	8	10	12	14	16	18	20
kVp	52	56	60	64	70	74	78	82	86	90	kVp	52	56	60	64	70	74	78	82	86	90
mA	100	100	100	100	100	100	100	100	100	100	mA	100	100	100	100	100	100	100	100	100	100
time	1/60	1/60	1/60	1/60	1/40	1/40	1/40	1/40	1/40	1/40	time	1/60	1/60	1/60	1/60	1/40	1/40	1/40	1/40	1/40	1/40
mAs	1.7	1.7	1.7	1.7	2.5	2.5	2.5	2.5	2.5	2.5	mAs	1.7	1.7	1.7	1.7	2.5	2.5	2.5	2.5	2.5	2.5
<b>Radius &amp; Ulna CC - 100 speed</b>					<b>Table Top</b>						<b>Table w/Grid</b>										
cm	2	4	6	8	10	12	14	16	18	20	cm	2	4	6	8	10	12	14	16	18	20
kVp	54	58	62	66	70	74	78	82	86	90	kVp	54	58	62	66	70	74	78	82	86	90
mA	100	100	100	100	100	100	100	100	100	100	mA	100	100	100	100	100	100	100	100	100	100
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3
<b>Radius &amp; Ulna Lateral - 100 speed</b>					<b>Table Top</b>						<b>Table w/Grid</b>										
cm	2	4	6	8	10	12	14	16	18	20	cm	2	4	6	8	10	12	14	16	18	20
kVp	54	58	62	66	70	74	78	82	86	90	kVp	54	58	62	66	70	74	78	82	86	90
mA	100	100	100	100	100	100	100	100	100	100	mA	100	100	100	100	100	100	100	100	100	100
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3
<b>Carpus &amp; Metacarpus DP - 100 speed</b>					<b>Table Top</b>						<b>Table w/Grid</b>										
cm	1	2	3	4	5	6	7	8	9	10	cm	1	2	3	4	5	6	7	8	9	10
kVp	60	62	64	66	68	70	72	74	76	78	kVp	60	62	64	66	68	70	72	74	76	78
mA	100	100	100	100	100	100	100	100	100	100	mA	100	100	100	100	100	100	100	100	100	100
time	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	time	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40
mAs	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	mAs	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5
<b>Carpus &amp; Metacarpus Lateral - 100 speed</b>					<b>Table Top</b>						<b>Table w/Grid</b>										
cm	1	2	3	4	5	6	7	8	9	10	cm	1	2	3	4	5	6	7	8	9	10
kVp	60	62	64	66	68	70	72	74	76	78	kVp	60	62	64	66	68	70	72	74	76	78
mA	100	100	100	100	100	100	100	100	100	100	mA	100	100	100	100	100	100	100	100	100	100
time	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	time	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40
mAs	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	mAs	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5
<b>Femur VD</b>			<b>Table Top</b>			<b>Table w/Grid</b>															
cm	6	8	10	12	14	16	18	20	22	24	cm	6	8	10	12	14	16	18	20	22	24
kVp	60	64	68	72	76	80	84	88	92	96	kVp	60	64	68	72	76	80	84	88	92	96
mA	100	100	100	100	100	100	100	100	100	100	mA	100	100	100	100	100	100	100	100	100	100
time	1/30	1/30	1/15	1/15	1/15	1/15	1/15	1/15	1/15	1/15	time	1/30	1/30	1/15	1/15	1/15	1/15	1/15	1/15	1/15	1/15
mAs	3.3	3.3	6.7	6.7	6.7	6.7	6.7	6.7	6.7	6.7	mAs	3.3	3.3	6.7	6.7	6.7	6.7	6.7	6.7	6.7	6.7

## InnoVet Technique Chart 60Hz

NOTE: 400 speed (except where noted) - automatic processing @ 94 degrees.

<b>Femur Lateral</b>		<b>Table Top</b>		<b>Table w/Grid</b>							
cm	<b>6</b>	<b>8</b>	<b>10</b>	<b>12</b>	<b>14</b>	<b>16</b>	<b>18</b>	<b>20</b>	<b>22</b>	<b>24</b>	
kVp	60	64	68	72	76	80	84	88	92	96	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/30	1/30	1/15	1/15	1/15	1/15	1/15	1/15	1/15	1/15	
mAs	3.3	3.3	6.7	6.7	6.7	6.7	6.7	6.7	6.7	6.7	
<b>Stifle CC - 100 speed</b>		<b>Table Top</b>				<b>Table w/Grid</b>					
cm	<b>2</b>	<b>4</b>	<b>6</b>	<b>8</b>	<b>10</b>	<b>12</b>	<b>14</b>	<b>16</b>	<b>18</b>	<b>20</b>	
kVp	52	56	60	64	68	72	76	80	84	88	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	
<b>Stifle Lateral - 100 speed</b>		<b>Table Top</b>				<b>Table w/Grid</b>					
cm	<b>2</b>	<b>4</b>	<b>6</b>	<b>8</b>	<b>10</b>	<b>12</b>	<b>14</b>	<b>16</b>	<b>18</b>	<b>20</b>	
kVp	52	56	60	64	68	72	76	80	84	88	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	
<b>Tibia &amp; Fibula CC - 100 speed</b>		<b>Table Top</b>				<b>Table w/Grid</b>					
cm	<b>2</b>	<b>4</b>	<b>6</b>	<b>8</b>	<b>10</b>	<b>12</b>	<b>14</b>	<b>16</b>	<b>18</b>	<b>20</b>	
kVp	52	56	60	64	68	72	76	80	84	88	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	
<b>Tibia &amp; Fibula Lateral - 100 speed</b>		<b>Table Top</b>				<b>Table w/Grid</b>					
cm	<b>2</b>	<b>4</b>	<b>6</b>	<b>8</b>	<b>10</b>	<b>12</b>	<b>14</b>	<b>16</b>	<b>18</b>	<b>20</b>	
kVp	52	56	60	64	68	72	76	80	84	88	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/60	1/60	1/60	1/60	1/30	1/30	1/30	1/30	1/30	1/30	
mAs	1.7	1.7	1.7	1.7	3.3	3.3	3.3	3.3	3.3	3.3	
<b>Tarsus &amp; Metatarsus DP - 100 speed</b>		<b>Table Top</b>									
cm	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	<b>8</b>	<b>9</b>	<b>10</b>	
kVp	60	62	64	66	68	70	72	74	76	78	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	
mAs	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	
<b>Tarsus &amp; Metatarsus Lateral - 100 speed</b>		<b>Table Top</b>									
cm	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	<b>8</b>	<b>9</b>	<b>10</b>	
kVp	60	62	64	66	68	70	72	74	76	78	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	1/40	
mAs	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	

## InnoVet Technique Chart 60Hz

NOTE: 400 speed (except where noted) - automatic processing @ 94 degrees.

<b>Birds VD</b>		<b>Table Top</b>									
cm	1	2	3	4	5	6	7	8	9	10	
kVp	50	51	52	52	53	53	54	54	55	56	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/20	1/20	1/15	1/15	1/15	1/15	1/15	1/15	
mAs	5	5	5	5	6.7	6.7	6.7	6.7	6.7	6.7	

  

<b>Birds Lateral</b>		<b>Table Top</b>									
cm	1	2	3	4	5	6	7	8	9	10	
kVp	50	51	52	52	53	53	54	54	55	56	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/20	1/20	1/15	1/15	1/15	1/15	1/15	1/15	
mAs	5	5	5	5	6.7	6.7	6.7	6.7	6.7	6.7	

  

<b>Pocket Pets DV</b>		<b>Table Top</b>									
cm	1	2	3	4	5	6	7	8	9	10	
kVp	53	54	55	56	57	58	59	60	61	63	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/15	1/15	
mAs	5	5	5	5	5	5	5	5	6.7	6.7	

  

<b>Pocket Pets Lateral</b>		<b>Table Top</b>									
cm	1	2	3	4	5	6	7	8	9	10	
kVp	53	54	55	56	57	58	59	60	61	63	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/15	1/15	
mAs	5	5	5	5	5	5	5	5	6.7	6.7	

  

<b>Reptiles DV</b>		<b>Table Top</b>									
cm	1	2	3	4	5	6	7	8	9	10	
kVp	53	54	55	56	57	58	59	60	61	63	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/15	1/15	
mAs	5	5	5	5	5	5	5	5	6.7	6.7	

  

<b>Reptiles Lateral</b>		<b>Table Top</b>									
cm	1	2	3	4	5	6	7	8	9	10	
kVp	53	54	55	56	57	58	59	60	61	63	
mA	100	100	100	100	100	100	100	100	100	100	
time	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/20	1/15	1/15	
mAs	5	5	5	5	5	5	5	5	6.7	6.7	